


Rev #	Revision Record	Date	Appvl
A	CHANGE STYLE OF WYE	6-4-98	KC
B	ADD WYE TEST CAP	1-8-01	KC
C	CHANGE TO OB PEDI WYE	6-22-01	KC
D	ADD 1/4" CUFF (INADVERTENT OMISSION)	3-14-03	NK
E	ADD 15 X 22mm ADAPTER (INADVERTENT OMISSION)	9-12-05	KC
F	REMOVE HOLDING ARM	2/14/10	NH



Extras included:

Adapter

24"

- All adapters depicted are 15mm X 22mm.
- All corrugated hose 24" in length.
- Not depicted (+):
- (1) 15mm X 15mm adapter joining two 24" lengths of hose
 - (1) accessory 15mm X 15mm adapter

Part Number: 40011

Hose ID: 15mm

SIGNATURES		DATE		BIO—MED DEVICES 61 Soundview Road Guilford, CT 06437 203/458—0202 FAX 458—0440		
dwg.		8—22—97				
chk'd						
MATERIAL:		NA				
TOLERANCES:		NA		TITLE		
				CIRCUIT, PEDIATRIC, DISPOSABLE		
				SCALE	DWG NO	REV
				none		
FINISH:		NA		FILE:	Pediatric	SHEET #1 OF 1 SHEETS