BIO-MED DEVICES MVP-10

Pediatric/Neonatal Ventilator















AVAILABLE FOR MRI

Pediatric/Neonatal

Ventilator

The MVP-10, with its ease of use and proven dependability, has remained a favorite of respiratory care professionals throughout its history. With its very wide range of I/E ratio settings and its ability to deliver a continuous flow of 21 - 100% oxygen, this is a very versatile ventilator.

- Continuous Flow
- Available as MRI compatible
- Portable, Compact and Light Weight
- May be used without interruption during transportation
- Gas powered: pressurized tank or wall outlet
- May be purchased as a "MRI Complete Package" shown here*

*Tanks not included





I/E Ratio variable, including inverse

O2 Concentration 21 - 100% Tidal Volume 0 - 666 ccCalibrated Flow Rate 0 - 20 LPM 0 - 70 cmH₂O Maximum Pressure Range Failsafe Maximum Pressure 80 cmH₂O PEEP/CPAP Range 0 - 18 cmH₂O

50+/-5 PSI gas source Power Supply

Logic Gas Consumption approx. 4 LPM (0.14 SCFM)

Flowmeter Ranges O₂: 0.6 to 10 LPM

Air: 0.6 to 10 LPM

System Pressure Gauge -10 to 120 cmH₂O

8"x9"x3" (20x23x7.4 cm) **Physical Dimensions** 5 lbs., 2 oz. (2.3 kg) Weight

32° to 122° F (0° to 50° C) Storage Temperature

Operating Temperature 14° to 122° F (-10° to 50° C)



Ordering Information

| | MVP-10 Catalog # | | | | |
|----------------------|--------------------------------|--|--|--|--|
| 2010 | MVP-10 Ventilator | | | | |
| 2010M | MVP-10 Ventilator/MRI | | | | |
| 2010MBC | MVP-10 MRI Complete Package | | | | |
| Standard Accessories | | | | | |
| 2030 | Disposable Infant Circuit | | | | |
| 2013 | 1" Pole Mount Bracket | | | | |
| 1010 | 10' O ₂ Supply Hose | | | | |
| 1011 | 10' Air Supply Hose | | | | |
| 1021 | Infant/Pediatric Test Lung | | | | |
| 2050 | Instruction Manual | | | | |

((





BIO-MED DEVICES, INC.

61 Sound View Road, Guilford, CT 06437 USA Voice: 1-800-2BIO MED (1-800-224-6633) (203) 458-0202 FAX: (203) 458-0440 Website: www.biomeddevices.com

| D | е | a | le | , |
|---|---|---|----|---|
| | | | | |